

ihcDirect® Pan-CK 4Abs

Anti-Human Pan-Cytokeratin (Clones AE1, AE3, C94, R226)

Abs: K31058-015, 150 tissue stains*

Kit: K41058-010, 100 tissue stains*

Kit: K41058-005, 50 tissue stains*

Intended Use: For In Vitro Diagnostic Use

Polymerized horseradish peroxidase (polyHRP)-labeled anti human pan-cytokeratin (Clones AE1, AE3, C94 and R226) antibody cocktail is intended for laboratory use to qualitatively identify by light microscopy the presence of cytokeratins which are expressed in tissues of epithelial origin, including squamous and glandular epithelial cells, in sections of formalin-fixed, paraffin-embedded (FFPE) and/or frozen tissues (FT) using immunohistochemistry (IHC) test method. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests and proper controls interpreted by a qualified pathologist and/or physician. This conjugate has been pre-diluted and optimized for IHC use without further dilution.

Summary and Explanation:

The ihcDirect® Pan-CK 4Abs is a polyHRP conjugated antibody cocktail which consists of equal amounts of three mouse monoclonal antibodies clones AE1, AE3 and C94 and one rabbit monoclonal antibody clone R226. Clones AE1 and AE3, always used together as an anti-pan-cytokeratin antibody cocktail, recognize cytokeratins 1-8, 10, 14 -16 and 19 in Western-Blot assay; clone C94 recognizes cytokeratins 8 and 18; clone R226 specifically recognizes CK5. The classic Pan-CK antibody cocktail AE1/AE3 has been the most popular Pan-CK cocktail in the market and is known to have limited reactivity in hepatocellular carcinomas (HCCs), renal cell carcinomas (RCCs), and pulmonary small cell carcinomas, and even on FFPE tissue specimens. Based on Novodiox test results on frozen tissues, this cocktail has limited reactivity in some lung cancer specimens and no to very poor reactivity with the cytokeratin proteins in epidermal layer of skin. This four-antibody cocktail is intended for laboratory use as a qualitative immunohistochemistry reagent for the identification of a broad range of cytokeratin proteins and for overcoming the poor reactivity of AE1/AE3 in frozen epithelial tissues, such as skin and lung.

Principle of Procedure:

The ready-to-use ihcDirect polyHRP Pan-CK 4Abs antibody conjugate cocktail is directly applied to pretreated tissue sections, where it binds to human cytokeratins. A DAB working solution is then applied to the tissue. The antibody-linked pHRP catalyzes the DAB to form a visible brown color product which precipitates at the site of the human Pan-CK location. The specimen may then be counterstained with hematoxylin and a coverslip applied. Results are viewed and interpreted using a light microscope. Volumes are based upon 100 µl antibody per tissue. This IHC reagent may be performed either manually or via automatic IHC staining system. Please read all procedures and Warnings and Precautions sections prior to use.

Reagents Provided:

Part No.	∇	Description
K31058-015	150*	15ml of ihcDirect Pan-CK 4Abs ready-to-use antibody conjugate for use with DAB Kit and ihc Blocker.
K41058-010	100*	10ml of ihcDirect Pan-CK 4Abs kit ready-to-use antibody conjugate, ihc Blocker and equivalent volume of ihc DAB and ihc DAB Diluent.
K41058-005	50*	5ml of ihcDirect Pan-CK 4Abs kit ready-to-use antibody conjugate, ihc Blocker and equivalent volume of ihc DAB and ihc DAB Diluent.

* At estimated volume of 100 µl of antibody conjugate per tissue

Immunogen	Clones	Species	Ig Class	Total Protein Conc.
Human Pan-CK	AE1, AE3, C94, R226	Mouse/Rabbit	IgG	10 mg/ml

Pan-CK 4Abs antibody is a cocktail of mouse and rabbit monoclonal antibodies to human cytokeratins. HRP is extracted from horseradish plant. This assay is designed

for stand-alone use as a kit K41058-### or as an individual antibody reagent (K31058-###) for use with the Novodiox ihc Blocker and DAB Kit sold separately.

The ihc DAB Diluent contains the buffered peroxide substrate. The ihc DAB (chromogen) contains 3,3'-diaminobenzidine which is dissolved in a proprietary buffer system with no hazardous chemicals at a reportable concentration. This reagent is light sensitive. For best results, minimize the time the vial is open.

Pan-CK 4Abs Kit Components for 5ml and 10ml Test Kits (K41058-###):

Kit Components	Component Part Numbers	Sizes
Pan-CK 4Abs pHRP	H31058-(R###) (005, 010)	5ml, 10ml
ihc Blocker	C30005-(###ML) (005, 010)	5ml, 10ml
ihc DAB Diluent	C30004-(###ML) (007, 013)	7ml, 13ml
ihc DAB	C30003-(###UL) (200, 375)	200µl, 375µl

Pan-CK 4Abs pHRP reagent components for 15ml volume (K31058-015 – Antibody Only):

Reagent Description	Component Part Numbers	Sizes
Pan-CK 4Abs pHRP	K31058-(R015)	15ml
For Use With Novodiox Antibodies:		
ihc Blocker Intl. USA	K51001-(015) (OR) K51002-(015)	15ml 15ml
DAB Kit	K50001-(###) (015, 030)	15ml, 30ml

Materials Needed but Not Provided:

The following reagents/supplies may be required in staining but are not provided:

- Frozen section fixative (ihc Fixative) or reagent grade acetone
- Positive and negative control tissues
- Microscope slides, positively charged (required)
- Staining jars, baths or processing tools
- ihc Wash Buffer (PBS-T)
- Pipettor and pipet tips
- Timer
- Antigen retrieval buffer (when using FFPE tissues)
- Peroxide blocker (optional)
- Instruments used for tissue pretreatment, such as water bath, or pressure cooker or microwave oven (when using FFPE tissues)
- Hematoxylin
- Xylene or Xylene substitute
- Ethanol
- Mounting medium
- Cover slips
- Light microscope (40 - 400x)

Novodiox Bulk Reagent Formulations:

- ihc Fixative, (375ml of methyl alcohol, 100ml of 37% formaldehyde and 25ml of glacial acetic acid).
- ihc Wash Buffer (PBS-T), (10 mM phosphate buffer, pH 7.2, 150 mM NaCl, 0.05% Tween-20).
- ihc Antigen Retrieval Buffer (10 mM Citric buffer, pH 6.0, 0.02% Tween 20).

Storage and Handling:

The reagent should be stored at 2-8°C. Do not freeze. This reagent is suitable for use until expiry date when stored at 2-8°C. Do not use the product after expiration date stamped on vial unless dating extension information is provided by Novodiox. If the reagent is stored under any conditions other than those specified in the package insert, they must be verified by the user. DAB working solution should be made just prior to use and is stable at 2-8°C during the day the reagents are made.

Specimen Preparation:

Paraffin Sections: Tissues routinely processed, neutral buffered 10% formalin-fixed are suitable for use prior to paraffin embedding. Consult references (Kiernan, 1981; Sheehan & Hrapchak, 1980). Variable results may occur as a result of prolonged fixation. Each section should be cut to the appropriate thickness



(approximately 4-5µm) and placed on a positively charged glass slide. Slides containing the tissue section may be baked for at least one hour but not exceeding 24 hours in a 58-60°C±5°C oven. Osseous tissues should be decalcified prior to tissue processing to facilitate tissue cutting and prevent damage to microtome blades (Kiernan, 1981; Sheehan & Hrapchak, 1980).

Frozen Tissue Sections: Frozen tissue is sectioned to the appropriate thickness (approximately 5 µm) and placed on a positively charged glass slide. Tissues should be fixed in either the NovodiAx iHC Fixative or reagent grade acetone for 30-seconds-to-1-minute immediately after sectioning. Reagent grade acetone may be kept cold, e.g. at cryostat temperatures or at room temperature. Following fixation, tissues may be stored in PBS-T for as long as a day.

Treatment of Tissues Prior to Staining: Pretreatment is tissue dependent and should be performed as suggested in the staining procedure sections.

Warnings and Precautions:

1. The Pan-CK 4Abs pHRP antibody conjugate is pre-diluted. Further dilution may reduce signal intensity or false-negative staining. These recommendations are for guidance only. Laboratory managers should determine their own procedures and quality policies.
2. Read and understand all of the NovodiAx Instructions for Use (IFUs) before product use.
3. For best results, minimize the time the iHC DAB vial is open. This reagent is light sensitive; store the mixed working solution of iHC DAB and iHC DAB Diluent away from light. Refrigerate when not in use.
4. Take reasonable precautions when handling reagents. Use protective equipment such as disposable gloves and lab coats when handling suspected carcinogens or toxic materials. Read Safety Data Sheets (SDS) prior to use.
 - a. Thimerosal is used as a preservative in iHC Blocker and the substance is classified as toxic substance. Inhalation causes respiratory and CNS effects and severe delayed neurotoxicity.
 - b. **WARNING!** DAB product contains 3,3'-diaminobenzidine which may cause genetic effects and/or cancer. If exposed or concerned, seek medical attention. See DAB SDS for more information.
5. Avoid contact of reagents with eyes and mucous membranes. If reagents come in contact with sensitive areas, wash with copious amounts of water.
6. Use charged slides to secure tissues appropriately.
7. Patient specimens and all materials that come into contact with patient specimens should be handled as bio-hazardous materials and disposed of appropriately.
8. Consult local or state authorities with regards to recommended methods of disposal of bio-hazardous and hazardous chemical waste materials.
9. Incubation time and temperature other than those specified may give erroneous results. The user must validate any such changes.
10. Use lab grade quality chemicals such as acetone, ethanol and water when preparing fixatives and buffers. Users should validate performance including stability for laboratory prepared reagents (at 1X).
11. Avoid microbial contamination of reagents.

Staining Procedures:

General Operating Notes:

1. Equilibrate all reagents to room temperature prior to use. Swirl or shake the pHRP-labeled antibody solution before use. **Do not vortex.** Calculate the amount of DAB working solution needed (100 µl per tissue) and **freshly** prepare DAB working solution by adding the ratio of 30 µl of iHC DAB to 1.0 ml of iHC DAB Diluent into an Eppendorf Tube.
2. It is best to prevent slides from drying out during the staining process to avoid unwanted background staining.
3. Gently and thoroughly wash tissues during manual wash steps. Avoid direct high velocity streams of wash that might tend to damage or cut delicate tissues.
4. Following each manual assay step, remove excess fluids on tissue slides with tissue paper. Excessive residual solution may dilute subsequent reagents, causing negative or uneven staining.
5. For the tissues with high level of oxidase activity, e.g. gastrointestinal or renal tissues, an additional blocking step with H₂O₂ is required to minimize background.
6. The following protocol has been validated at temperatures between 21°-30°C (70°-86°F) for incubating iHC Blocker, Pan-CK 4Abs pHRP and DAB working

solution. If room temperature is less than 21°C, incubate labeled antibody for a longer period of time (≥4 minutes depending upon temperature). Consistent results have been obtained using a slide warmer set to 30°C.

Frozen Tissue Sections:

1. Place frozen tissue sections into fixation solution immediately after sectioning. **Prolonged exposure to room or freezing temperatures may alter targeted epitopes.**
2. The DAB working solution incubation step is a range from 1-3 minutes. Users should determine the optimal incubation time for their lab environment and observe the brown color formation via visual inspection during incubation.

Test Timing Est. (10-minute IHC protocol for frozen tissue sections):

Procedure	Time in minutes
Frozen	
Fixation, use Acetone or iHC Fixative	Start
- Wash with iHC Wash Buffer	0:15
Block with iHC Blocker	1
- Tap and Blot to remove excess blocker	- - -
Pan-CK 4Abs pHRP	3
- Wash with iHC Wash Buffer	0:15
- Tap and Blot to remove excess wash buffer	- - -
DAB working solution	1-3
- Wash with iHC Wash Buffer	0:15
Hematoxylin counterstain	0:20
- Wash with iHC Wash Buffer	0:15
Mounting media and coverslip	0:45
Total	10

Paraffin Tissues:

1. Deparaffinization: Soak slides in Xylene 3 times for 5 minutes each. Next, 3 minutes each in 100%, 95% and 75% ethanol. Then wash slides with tap water in slide tank for two times, 2 minutes each time.
2. Antigen retrieval: Using a water bath, incubate slides in antigen retrieval buffer in a slide tank at 96°C for 30 minutes, then cool the slides down to room temperature for 30 minutes. Rinse the slides twice with tap water, 2 minutes each time.
3. (Optional) Block tissue with H₂O₂: Soak the slides in 3% H₂O₂ in slide tank, stand for 10 minutes. Rinse the slides with tap water twice and then wash once with PBS-T in slide tank for 2 minutes.
4. Dispense 100 µl of iHC Blocker covering the entire tissue and incubate at room temperature for 15 minutes. Remove iHC Blocker as much as possible but do not rinse the slides with PBS-T or water.
5. Dispense 100µl of pHRP labeled anti-human Pan-CK 4Abs antibody cocktail on slides covering the entire tissue and incubate for 15-30 minutes at room temperature. Rinse the slides three times with PBS-T in slide tank, 2 minutes each time. Note: Place slides in a wet chamber to prevent evaporation if longer incubation times are used.
6. Dispense 100µl of DAB working solution covering the entire tissue, incubate for 3-10 minutes at room temperature. Rinse the slides twice with tap water in slide tank, 2 minutes each time.
7. Counterstaining: Add hematoxylin and incubate for 1 minute at room temperature. Rinse twice with tap water for 2 minutes, each time.
8. Dehydration: Soak slides in the following order: 75% ethanol for 3 minutes, 95% ethanol for 3 minutes, 100% ethanol for 3 minutes and Xylene twice at 5 minutes each time.
9. Applying Coverslip: Add one drop of permanent mounting medium on both the slide and the coverslip, then apply coverslip.



Quality Control Procedures:

Positive and negative controls should be run simultaneously with patient specimens.

Positive Tissue Control: The recommended positive control tissues for this antibody are known Pan-CK 4Abs positive tissues. One positive tissue control for each set of test conditions should be included in each staining run. Previous tissue specimen and in some cases, an individual's own tissue may be used as a control.

The tissues used for the positive control should be selected from patient specimens with well-characterized low levels of the positive target activity that gives weak positive staining. Known positive tissue controls such as breast, lung and colon should only be utilized for monitoring the correct performance of processed tissues and test reagents, rather than as an aid in formulating a specific diagnosis of patient samples. If the positive tissue controls fail to demonstrate positive staining, results with the patient specimens should be considered invalid.

Negative Tissue Control: The same tissue used for the positive control may be used as the negative tissue control. The variety of cell types in most tissue sections offers internal negative control sites. But this should be verified by the user. The components that do not stain should demonstrate the absence of specific staining and provide an indication of non-specific background staining. If specific staining (false positive staining) occurs in the negative tissue control, results with the patient specimens should be considered invalid.

Troubleshooting:

If an unexpected staining pattern occurs on control tissues or patient samples, consider the following:

1. No staining: If no staining on positive control slide, please verify whether (1) chromogen was prepared freshly and correctly, (2) reagents were used in the specified order, (3) pHRP-labeled antibody was indeed added, and (4) for FFPE tissue, dewaxing and antigen retrieval were performed inadequately. Perform any corrective actions required and then repeat the procedure.
2. Weak staining: Please check whether (1) the reagents have expired, (2) room temperature was below 21°C if a 30°C slide warmer was not used, (3) chromogen was prepared freshly, (4) too much washing solution remained on slide and diluted next reagent, and (5) for FFPE tissue, dewaxing and antigen retrieval were performed inadequately. Perform any required corrective actions and repeat the procedure.
3. High background: Possible causes include (1) insufficient washing, (2) blocker not applied or washed out after application, (3) specimens dried out, (4) prolonged incubation with chromogen, (5) prolonged incubation with pHRP-labeled antibody and (6) specimens contain high level of endogenous peroxidase and need an additional blocking step (refer to the Block tissue with H₂O₂ step in "Staining Procedures Paraffin Tissues"). Perform any required corrective actions and repeat the procedure.
4. The ihc DAB chromogen volume provided is based on a typical user. Occasionally, materials can stick to either the lid or the side of the vial. To gain access to all of the material, it may be necessary to centrifuge at a slow speed or tap down the bottle using caution prior to use.

If an unexpected staining pattern is observed on control tissues or patient samples which cannot be explained by variations in laboratory procedures or a problem with the antibody is suspected, contact NovodiAx Technical Support or your local distributor immediately. Within the US and Canada call 1 (888) 439-2716 ext. 2 or 1 (510) 342-3043 ext. 2.

Expected Results:

Intense brown color stains with a clean background if Pan-CK-expression cells exist. No brown color stains if no Pan-CK-expression cells exist. Interpretation of the staining result is solely the responsibility of the user.

General Limitations:
















Immunohistochemistry is a multistep diagnostic process that requires specialized training in the selection of the appropriate reagents; tissue selection, fixation, and processing; preparation of the IHC slide; and interpretation of the staining results. Improper fixation, freezing, thawing, washing, drying, heating, sectioning or contamination with other tissues or fluids may produce artifacts, antibody trapping, or false negative results. Inconsistent results may be due to variations in fixation and

embedding methods, or to inherent irregularities within the tissue (Nadji M, Morales AR. 1983).

The manufacturer provides these antibodies/reagents at optimal dilution for use following the provided instructions for IHC on prepared tissue sections. Any deviation from recommended test procedures may invalidate declared expected results; appropriate controls must be employed and documented. Users who deviate from recommended test procedures must accept responsibility for interpretation of patient results under these circumstances.

Performance Characteristics:

The ihcDirect Pan-CK 4Abs pHRP test performance has been determined using both frozen and FFPE tissue sections. NovodiAx has conducted studies to evaluate the performance of the antibody conjugate, separate NovodiAx reagents and ancillary supplies. The antibodies and systems have been found to be sensitive and show specific binding to the antigen of interest with minimal to no binding of non-specific tissues or cells. The NovodiAx antibody and separate reagents have shown reproducible and consistent results when used within a single run, between runs and between lots. These products have been determined to be stable for the periods of time specified on the labels either by standard real-time and/or accelerated methods. NovodiAx ensures product quality by testing each lot of material and by testing materials at regular intervals and via surveillance programs.

ihcDirect Pan-CK 4Abs Key to Symbols			
	<i>In vitro</i> diagnostic medical device		pHRP Pan-CK 4Abs antibody conjugate
	Catalog Number		Blocking reagent
	Use by: YYYY-MM-DD		DAB Chromogen reagent
	Consult Instruction for Use		DAB Diluent reagent
	Batch Code		Health Hazard
	Contains sufficient for < n > tests		Manufacturer
	Temperature Limitation		CE Mark
	Authorized European Representative		

Instructions for Use (IFU) Access:

To obtain a translation or the latest electronic version of an IFU document, visit our website at <https://www.novodiox.com/support/literature/> (ihcDirect IFU). Printed copies of an IFU document may be obtained by contacting NovodiAx Technical Support or your local distributor.

Bibliography:

1. Kiernan JA. Histological and Histochemical Methods: Theory and Practice. New York: Pergamon Press 1981.
2. Sheehan DC and Hrapchak BB. Theory and Practice of Histotechnology. St. Louis: C.V.Mosby Co. 1980.
3. Nadji M, Morales AR. Immunoperoxidase, part I: the techniques and its pitfalls. Lab Med, 1983; 14:767-771.
4. Woodcock-Mitchell J, Eichner R, et al. Immunolocalization of Keratin Polypeptides in Human Epidermis Using Monoclonal Antibodies. The Journal of Cell Biology. November 1982; 95:580-588.
5. Weiss RA, Eichner R, et al. Monoclonal Antibody Analysis of Keratin Expression in Epidermal Diseases: A 48- and 56-kdalton Keratin as Molecular Markers for Hyperproliferative Keratinocytes. The Journal of Cell Biology. April 1984; 98:1397-1406.
6. Moll R, Divo M, et al. The human keratins: biology and pathology. Histochem Cell Biol. 2008; 129:705-733.
7. Ordenez NG. What are the current best immunohistochemical markers for the diagnosis of epithelioid mesothelioma? A review and update. Human Pathology. 2007; 38:1-16.

